

Test Facility:
 BioGenius GmbH
 TechnologiePark
 Campus 1
 Friedrich-Ebert-Straße 75
 51429 Bergisch Gladbach, Germany

No. CoA: AQ23-065

Certificate of Analysis

Study No. : AQ23-065
Sponsor : CO medizell GmbH
Test Item : Trionic™
Batch No. : 202167
Production Date : 2020-11
Date of Analysis : 2023-12-13 (AQ724) and 2023-12-28 (AQ725)
Test Method : AQ724 (HPLC Determination of Benzalkoniumchloride in Formulations) and AQ725 (HPLC-MS-Determination of DDAC)
Retest Date : 2024-12

	Specification	Result [%]	Requirement [%]	RSD [%]
1.1	Content (HPLC-MS-ESTD)			
1.2	DDAC (N,N-Didecyl-N,N-dimethylammoniumchloride)	9.74	8.35 – 10.21	1.11
2.1	Content (HPLC-ESTD)			
2.2	BAC (Benzalkoniumchloride)	10.19	9.18 – 11.22	0.23

This analysis was conducted in a Good Laboratory Practice (GLP) certificated facility.

Study Director:

2024-01-15
Date

Th. Röper
Theresa Röper (M.Sc.)